



## Vitamin E

### 1. Analysis Condition

#### *YL 9100 Series Vitamin Analyzer*

Mobile Phase : Methanol

Flow Rate : 1.0ml/min

Column Oven : 35°C

Detector : UV/Vis 280nm

Column : C18 (4.6\*250mm, 5um)

Injection Volumn : 20ul sample loop

### 2. Chromatogram

