



Fat Soluble Vitamins

1. Analysis Condition

YL9100 Series Vitamin Analyzer
Mobile Phase : ACN:MeOH = 75:25
Flow Rate : 1.0ml/min
Column Oven : 35°C
Detector : UV/Vis 280nm
Column : C18 (4.6*250mm, 5um)
Injection Volume : 20ul sample loop

2. Chromatogram

