



## Fat Soluble Vitamins

### 1. Analysis Condition

#### **YL9100 Series Vitamin Analyzer**

Mobile Phase : MeOH : Water = 95:5

Flow Rate : 1.0ml/min

Column Oven : 35°C

Detector : UV/Vis 280nm

Column : C18 (4.6\*150mm, 5um)

Injection Volume : 20ul sample loop

### 2. Chromatogram

