



Vitamin D₂ in Tablet

1. Analysis Condition

YL9100 Series Vitamin Analyzer

Column : C18 (1.5X250mm, 5 μ m)

Detector : UV/Vis Wavelength : 280nm

Mobile phase : ACN:MeOH=75:25

Flow rate : 0.2 ml/min

Injection volume : 30 μ l

2. Chromatogram

